



Industrie Service

Choose certainty.
Add value.

CONFIRMATION

on the

type test of a flue gas analyser

Date: 2016-04-11

Our reference:
IS-TAF-MUC/thi

Order no. 2397956

Document:
ME11840016_BST_eng.docx

Test Laboratory: TÜV SÜD Industrie Service GmbH
Abteilung Feuerungs- und Wärmetechnik
Prüfbereich messtechnische Einrichtungen

Page 1

Subject of Test: Flue gas analyser type **DELTA**smart

Equipment tested functional modules for the measurement of

- O₂ 0 to 21,0 Vol. %
- CO 0 to 2000 ppm
- flue gas temperature 0 to 400°C
- inlet air temperature 0 to 50°C
- pressure (draught) -40 to 190 Pa
- pressure (differential) 0 to 10000 Pa

The document consists of
1 pages

Excerpts from this document
may only be reproduced and
used for advertising purposes
with the express written
approval of TÜV SÜD Industrie
Service GmbH.

Ordering Company: MRU GmbH
Fuchshalde 8
D - 74172 Neckarsulm-Obereisesheim

The test results refer exclusively
to the units under test.

Basis of Test: DIN EN 50379-1: 2013-06
DIN EN 50379-2: 2013-06

Test Report: No. M-E 1184-00/16 dated 2016-04-11

The tests have been performed with positive results.
The results in detail, the evaluation of the results and the conclusions out of
the results are described in the above mentioned test report.

Feuerungs- und Wärmetechnik

Johannes Steiglechner



Headquarters: Munich
Trade Register Munich HRB 96 869
VAT ID No. DE129484218
Information pursuant to Section 2(1)
DL-InfoV (Germany) at
www.tuev-sued.com/imprint

Supervisory Board:
Karsten Xander (Chairman)
Board of Management:
Ferdinand Neuwieser (CEO),
Dr. Ulrich Klotz, Thomas Kainz

Telefon: +49 89 51 90 - 1027
Telefax: +49 89 51 90 - 3307
E-mail feuerung@tuev-sued.de
www.tuev-sued.de/is

TÜV®

TÜV SÜD Industrie Service GmbH
Feuerungs- und Wärmetechnik
Ridlerstrasse 65
80339 Munich
Germany