

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

MGAprime Q

manufactured by:

MRU GmbH

Fuchshalde 8
Neckarsulm, 74172
Germany

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

Environment Agency Guidance
“MCERTS for stack emissions monitoring equipment at industrial installations”
- Transportable Continuous Emissions Monitoring Systems(T-CEMS)
Published 20 October 2020
EN 15267-1:2009, EN 15267-2:2009, EN 15267-4:2017
QAL 1

Certification ranges:

Supplementary ranges:

CO	0 - 220 mg/m ³	0 - 3,750 mg/m ³
CO ₂	0 - 20 vol. %	
NO	0 - 270 mg/m ³	0 - 2,680 mg/m ³
NO ₂	0 - 308 mg/m ³	0 - 1,025 mg/m ³
O ₂	0 - 25 vol. %	

Project number: 80049732
Certificate number: Sira MC200366/00
Initial certification: 22 December 2020
This certificate issued: 22 December 2020
Renewal date: 21 December 2025



Andrew Young
Environmental Team Manager

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service



Unit 6, Hawarden Industrial Park
Hawarden, Deeside, CH5 3US
Tel: +44 (0)1244 670 900



*The MCERTS certificate consists of this document in its entirety.
For conditions of use, please consider all the information within.
This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*